

Nitrate – NED / Zinc Reduction Method

Version 1 / April 2026

Applications and Industries

Freshwater, Seawater and Industrial water

References

- Nelson J. L., Kurtz, L. T., and R. H. Bray Rapid Determination of Nitrates and Nitrites. Analytical Chem., V26, p 1081-2 (1954)
- APHA Standard Methods, 23rd ed., Method 4500-NO₂ - B - 2000.
- EPA Methods for Chemical Analysis of Water and Wastes, Method 354.1 (1983).

Chemistry

Nitrate is reduced to nitrite by shaking the sample with a zinc metal catalyst and a citrate buffer. The nitrite is reacted with sulfanilic acid and N-(1-naphthyl)ethylenediamine dihydrochloride (NED) in an acidic solution to produce a intense reddish-purple azo dye. Results are expressed as ppm (mg/L) nitrate-nitrogen (NO₃-N). This method is applicable to industrial wastewaters, drinking, and surface waters. These Nitrate in water test kits can also be used for the analysis of seawater. To convert results from ppm NO₃-N to ppm NO₃, multiply by 4.43.

Available Analysis Systems

Visual colorimetric: CHEMets®, HR CHEMets®

Instrumental colorimetric: Vacu-vials®

Storage Requirements

Products should be stored in the dark at room temperature.

Correction for Nitrite

This method will measure nitrate in the presence of low levels of nitrite:

Corrected Nitrate, ppm NO₃-N =

Total (K-6973, ppm NO₃-N) - 1.14 × Nitrite(K-7013, ppm NO₂-N)

Note: K-7013 NO₂-N is equivalent to the K-6973 test without the addition of the Zinc powder and S-6902 (buffer), but with 5 drops of S-7004 (acidifier).

Interference Information

- Cupric copper [Cu(II)] – Negative bias above 0.3 mg/L Cu(II).
- Ferrous Iron [Fe(II)] - 1% bias/mg/L Fe(II)
- Alkalinity – Negative bias above 500 mg/L CaCO₃ alkalinity, adjust the sample to pH 6 with HCl.
- For sample color or turbidity use CHEMetrics' Sample Zeroing Accessory Pack.
- Chromate interferes by causing the formation of a yellow color. **CHEMetrics' Sample Zeroing Accessory Pack (A-0025)** can be used to correct for potential errors during instrumental analysis.

Accuracy Statement

Statements of accuracy are based on laboratory tests performed under ideal testing conditions using standards of known concentration prepared in deionized water.

Visual colorimetric:

CHEMets® kits: ± 1 color standard increment

Instrumental colorimetric

Vacu-vials® kit, K-6973:

≤ 0.04 ppm at 0 ppm NO₃-N

± 0.03 ppm at 0.10 ppm NO₃-N

± 0.20 ppm at 1.125 ppm NO₃-N

Vacu-vials® kit, K-6983 (High Range 10X):

≤ 0.40 ppm at 0 ppm NO₃-N

± 0.3 ppm at 1.0 ppm NO₃-N

± 2.0 ppm at 11.25 ppm NO₃-N

Safety Information

Safety Data Sheets (SDS) are available upon request and at www.sdsfetch.com. Read SDS before using these products. Breaking the tip of an ampoule in air rather than water may cause the glass ampoule to shatter. Wear safety glasses and protective gloves.